

Borosil Glass Works Limited

CIN: L99999MH1962PLC012538

Corporate Office: 1101, Crescenzo, G-Block, Opp. MCA Club,
Bandra Kurla Complex, Bandra (E) Mumbai - 400 051, India
Tel.: +91 22 6740 6300, Fax: +91 22 6740 6514, Email: borosil@borosil.com



www.borosil.com

Class "A" TEST CERTIFICATE

Batch Certificate

Certificate No. : 04-15-399 Date : 17-04-15
Test Method : Gravimetric Method.
Item Tested : Volumetric Flask
Catalogue No. : 5653
Batch No. : 04-15
Item Tested On : 17-04-15

The Product was found to contain the following volume at 20 °C

Certified Volume : 100 ml
Tolerance : ± 0.100 ml DT : N/A
Stan. Deviation : ± 0.013757 ml
Measuring uncertainty : ± 0.003816 ml

The product has been calibrated with following devices :

Using Balance No. : BAL16 Make : Sartorius
Thermometer No. : TH024 Range : 0-50 °C

All the above devices have valid calibration traceable to NPL, India; traceable to International Standards.

The above product have been examine for compliance to -Class A.

It has been found to satisfy the specified requirements. The above item has been calibrated at 24.6 °C at RH 39%.

The above calibration results are valid at the time of calibration under lab condition. The report shall not be reproduced, except in full, unless with written permission from M/s Borosil Glass Works Ltd.

Calibrated By : SAGAR
Date of Issue : 17-04-15

Authorised Signatory

V. Rama Sankar

Whole Time Director
Borosil Glass Works Ltd.

Regd. office : Khanna Construction House, 44, Dr. R.G.Thadani Marg, Worli, Mumbai - 400 018. Tel. : +91 22 2493 0362